



A REVIEW ON HPLC METHOD DEVELOPMENT AND VALIDATION

Manjiri Shinde¹, Dr. Satish Kumar², Dr. Arunabha Mallik³, Dr. N Jyothi⁴

^{1,2,3,4}Department of Pharmaceutical Analysis, MLR Institute of Pharmacy, Dundigal, Hyderabad, Telangana-500043

ABSTRACT

HPLC is the dominant separation technique to detect, separate and quantify the drug. A number of chromatographic parameters were analyzed to optimize the method like sample pretreatment, choosing mobile phase, column, detector selection. The objective of this article is to review the method development, optimization and validation. HPLC method development depends on chemical structure of the molecules, synthetic route, solubility, polarity, pH and pKa values, and functional groups activity etc. Validation of HPLC method as per ICH Guidelines gives information regarding various stages and knowing characteristics like Accuracy, specificity, linearity limit of detection, limit of quantification.

KEYWORDS: High Pressure Liquid Chromatography (HPLC), Method validation, Method development

1. INTRODUCTION

High Performance Liquid Chromatography is now one of the most powerful tools in analytical chemistry. It has the ability to separate, identify, and quantify the compounds that are present in any sample that can be dissolved in a liquid. High performance liquid chromatography (HPLC) is the most accurate analytical methods widely used for the quantitative as well as qualitative analysis of drug product. The principle is that a solution of the sample is injected into a column of a porous material (stationary phase) and a liquid (mobile phase) is pumped at high pressure through the column. The separation of sample is based on the differences in the rates of migration through the column arising from different partition of the sample between the stationary and mobile phase. Depending upon the partition behaviour of different components, elution at different time takes place. The sample compound with the greater affinity to the stationary layer will travel slower and for a shorter distance in comparison to compounds with less affinity which travel faster and for a longer distance. The High Performance Liquid Chromatography is more versatile than gas chromatography since (a) it is not limited to

volatile and thermally stable samples, and (b) the choice of mobile and stationary phases is wider.

HPLC has numerous advantages like

- Simultaneous Analysis
- High Resolution
- High Sensitivity
- Good repeatability
- Small sample size
- Moderate analysis condition.

Classification of HPLC can be done as

- preparative HPLC and analytical HPLC (based on scale of operation)
- affinity chromatography, adsorption chromatography, size exclusion chromatography, ion exchange chromatography, chiral phase chromatography (based on principle of separation)
- gradient separation and isocratic separation, (based on elution technique)
- normal phase chromatography and reverse phase chromatography (based on modes of operation).

**METHOD DEVELOPMENT ON HPLC**

A step involved in method development of HPLC is as follows:

- 1 Understanding the Physicochemical properties of drug molecule.
2. Selection of chromatographic conditions.
- 3 Developing the approach of analysis.
- 4 Method optimization
- 5 Method validation

1. Understanding the Physicochemical properties of drug molecule.

Physicochemical properties of a drug molecule play an important role in method development. For Method development one has to study the physical properties like solubility, polarity, pKa and pH of the drug molecule. Polarity is a physical property of a compound. It helps an analyst, to decide the solvent and composition of the mobile phase. The solubility of molecules can be explained on the basis of the polarity of molecules. Polar, e.g. water, and nonpolar, e.g. benzene, solvents do not mix. In general, like dissolves like i.e., materials with similar polarity are soluble in each other. The selection of mobile phase or diluents is based on the solubility of analyte. The analyte must be soluble in diluents and must not react with any of its component. pH and pKa plays an important role in HPLC method development.

1. Selection of chromatographic conditions:

Selection of column: Selection of the stationary phase/column is the first and the most important step in method development. The development of a rugged and reproducible method is impossible without the availability of a stable, high performance column. To avoid problems from irreproducible sample retention

during method development, it is important that columns be stable and reproducible.

Buffer Selection

Choice of buffer is governed by the pH that is desired. The typical pH range for reversed phase on silica based packing is pH 2 to 8. It is important that the buffer has a pKa close to the desired pH since buffer controls pH best at their pKa. A rule is to choose a buffer with a pKa value <2 units of the desired mobile phase pH.

Buffer Concentration

Generally, a buffer concentration of 10-50 mM is adequate for small molecules. Generally, no more than 50% organic should be used with a buffer. This will depend on the specific buffer as well as its concentration. Phosphoric acid and its sodium or potassium salts are the most common buffer systems for reversed-phase HPLC. Sulfonate buffers can replace phosphonate buffers when analyzing organophosphate compounds.

Selection of Mobile Phase: The mobile phase effects resolution, selectivity and efficiency. Mobile phase composition (or solvent strength) plays an important role in RP-HPLC separation. Acetonitrile (ACN), methanol (MeOH) and tetrahydrofuran (THF) are commonly used solvents in RP-HPLC having low UV cut-off of 190, 205 and 212nm respectively. These solvents are miscible with water. Mixture of acetonitrile and water is the best initial choice for the mobile phase during method development.

Selection of Detectors

Detector	Type of compound can be detected
UV-Visible & Photodiode array	Compounds with chromophores, such as aromatic rings or multiple alternating double bonds.
Fluorescence detector	Fluorescent compounds, usually with fused rings or highly conjugated planer system.
Conductivity detector	Charged compounds, such as inorganic ions and organic acid.
Electrochemical detector	For easily oxidized compounds like quinines or amines
Refractive Index detector & Evaporative light scattering detector	Compounds that do not show characteristics usable by the other detectors, eg. polymers, saccharides.

3. Developing the approach for analysis

While developing the analytical method on RP-HPLC the first step which is followed is the selections of various chromatographic parameters like

selection of mobile phase, selection of column, selection of flow rate of mobile phase, selection of pH of mobile phase. All of these parameters are selected on the basis of trials and followed by considering the



system suitability parameters.

4. Sample preparation

Sample preparation is an essential part of HPLC analysis, intended to provide a reproducible and homogenous solution that is suitable for injection onto the column. The aim of sample preparation is a sample aliquot that, is relatively free of interferences, will not damage the column, and is compatible with the intended HPLC method that is, the sample solvent will dissolve in the mobile phase without affecting sample retention or resolution. Sample preparation begins at the point of collection, extends to sample injection onto the HPLC column.

5. Method optimization

Identify the “weaknesses” of the method and optimize the method through experimental design. Understand the method performance with different conditions, different instrument set ups and different samples.

6. Method Validation

Validation is the confirmation by examination and the provision of objective evidence that the particular requirements for a specific intended use are fulfilled. A process of evaluating method performance and demonstrating that it meets a particular requirement. In essence, it knows what your method is capable of delivering, particularly at low concentrations.

Types of Analytical Procedures to be validated

The discussion of the validation of analytical procedures is directed to the four most common types of analytical procedures:

- Identification tests;
- Quantitative tests for impurities' content;
- Limit tests for the control of impurities;

Components of method validation: The following are typical analytical performance characteristics which may be tested during methods validation:

1. Accuracy
2. Precision
3. Linearity
4. Detection limit
5. Quantitation limit
6. Specificity
7. Range
8. Robustness

1) Accuracy

Accuracy is the closeness of understanding between the worth which is acknowledged either as an ordinary genuine quality or an acknowledged reference quality, and the quality discovered.

2) Precision

It communicates closeness of understanding (level of diffuse) between a progression of estimations acquired from various testing of the same homogeneous specimen under the recommended conditions. Accuracy may be considered at three levels: repeatability, transitional exactness and reproducibility.

3) Linearity

The linearity of a systematic strategy is its capacity (inside an offered extent) to acquire test outcomes that are specifically relative to the convergence of analyte in the specimen. Linearity is dictated by a progression of three to six infusions of five or more norms.

4) Detection limit

The detection limit of an individual explanatory method is the most minimal measure of analyte in an example which can be recognized yet not so much quantitated as an accurate quality.

5) Quantitation limit

The quantitation limit of an individual expository system is the least measure of analyte in an example which can be quantitatively decided with suitable accuracy and exactness. The quantitation limit is a parameter of quantitative tests for low levels of mixes in test lattices, and is utilized especially for the determination of polluting influences and/or corruption items

6) Specificity

Selectivity and specificity are once in a while utilized conversely to depict the same idea in strategy approval. Specificity is the capacity to evaluate unequivocally the analyte in the vicinity of parts that may be required to be exhibit. The specificity of a test system is controlled by contrasting test results from an investigation of tests containing contaminations, debasement items, or placebo fixings with those got from an examination of tests without debasements, corruption items, or placebo fixing.

**7) Range**

The Range is the interim between the upper and lower convergence of analyte in the example (counting these focuses) for which it has been shown that the systematic method has a suitable level of exactness, precision and linearity

8) Robustness

The robustness of an analytical procedure is a measure of its capacity to remain unaffected by small, but deliberate, variations in method parameters and provides an indication of its reliability during normal usage

CONCLUSION

This review describes about RP-HPLC Technique. The method development and validation are continuous and interrelated processes that measure a parameter as intended and establish the performance limits of the measurement. The selection of Column, buffer, detector and wavelength and another conditions composition (organic and pH) plays a dramatic role on the separation selectivity The advantages of HPLC technique were high selectivity, sensitivity, economic, less time consuming and low limit of detection. Final optimization can be performed by changing the gradient slope, temperature and flow rate as well as the type and concentration of mobile-phase modifiers. Optimized method is validated with various parameters (e.g. specificity, precision, accuracy, detection limit, linearity, etc.) as per ICH guidelines.

REFERENCES

1. Rao BV, Sowjanya GN, Ajitha A, Rao Uma MV. A review on stability indicating HPLC method development, *World journal of pharmacy and pharmaceutical sciences*. 2015; 4(8):405-423.
2. Rajan HV. Development and validation of HPLC method - A Review. *International Journal of current research in pharmacy*. 2015; 1(2):55-68.
3. Kumar V, Bharadwaj R, Gupta G, Kumar S. An Overview on HPLC Method Development, Optimization and Validation process for drug analysis. *The Pharmaceutical and Chemical Journal*. 2015; 2(2):30-40.
4. Gupta V, Jain AD, Gill NS, Gupta K. Development and validation of HPLC method - a review. *International Research Journal of Pharmaceutical and Applied Sciences*. 2012; 2(4):17-25.
5. Sonia K, Nappinnai M. Development and validation of HPLC and UV-visible spectrophotometric method for the pharmaceutical dosage form and biological fluid -review. *European Journal of Biomedical and Pharmaceutical sciences*. 2016; 3(3): 382-391.
6. Sánchez MLF. Chromatographic techniques, European RTN Project, GLADNET, retrieved on 05-09-2013.
7. Snyder LR, Kirkland JJ, Glach JL. *Practical HPLC Method Development*, John Wiley and Sons, New York, 1997; 158-192.
8. HPLC – Chemiguide. May 2, 2007. www.chemiguide.co.uk
9. Rao G, Goyal A. An Overview on Analytical Method Development and Validation by Using HPLC. *The Pharmaceutical and Chemical Journal*, 2016; 3(2): 280-289.
10. Mcpolin Oona. *An Introduction to HPLC for Pharmaceutical Analysis*. Mourn Training Service. 11-12.
11. <http://www.scribd.com/doc/9508765/Physical-Properties-of-Drug>.
12. Buffers and pH Buffers: available from: www.xtremepapers.com.
13. Charde MS, Welankiwar AS and Kumar J. Method development by liquid chromatography with validation. *International Journal of Pharmaceutical Chemistry*. 2014; 4(2):57-61.
14. Ranjit Singh. HPLC method development and validation. *J Pharm Educ Res* 2013; 4(1): 26-33.
15. Sabir AM, Molony M, Parminder SB. HPLC Method Development and validation: A Review. *International research Journal of pharmacy*. 2013; 4(4):39-46.
16. Suresh Babu VV et.al. Validated HPLC Method for Determining Related Substances in Compatibility Studies and Novel Extended Release Formulation for Ranolazine. *J Chromatograph Separat Techniq*. 2014; 5:209.
17. Gengaihi SEI et.al. Antioxidant Activity of Phenolic Compounds from Different Grape Wastes. *J Food Process Technol*. 2014; 5:296.
18. Shanmugam R et.al. Bioanalytical Method Development and Validation for Herbal Quercetin in Nano Formulation by RP-UFLC in Rabbit Plasma. *J Bioequiv Availab*. 2013; 5: 191-196
19. Bais S et.al. Method Development and Validation for Desogestrel and Ethinylestradiol in Combined Pharmaceutical Dosage Form by RP-HPLC. *Pharm Anal Acta*. 2013; 4: 262.
20. Tengli AR and Gurupadayya BM Method Development and Validation of Tablet Dosage form Containing Losartan, Atenolol and Hydrochlorthiazide Using Internal Standard by RP-HPLC. *J Chromat Separation Techniq*. 2013; 4: 180.
21. Tengli AR et.al. Method Development and Validation of Metformine, Pioglitazone and Glibenclamide in Tablet Dosage Form by using RP-HPLC. *Biochem Anal Biochem*. 2013; 2: 130.
22. de Figueiredo NB et.al. Determination of 3,4-



- methylenedioxyamphetamine (MDMA) in Confiscated Tablets by High-Performance Liquid Chromatography (HPLC) with Diode Array Detector. J Forensic Res.2010; 1:106.*
23. Pedro AQ et al. An Improved HPLC Method for Quantification of Metanephrine with Coulometric Detection. *J Chromatograph Separat Techniq. 2014; 5:217. RRJPPS | Volume 4 | Issue 1 | January - March, 2015 48 e-ISSN: 2320-1215 p-ISSN: 2322-0112*
24. Abdallah MA Validated Stability-indicating HPLC and Thin Layer Densitometric Methods for the Determination of Pazufloxacin: Application to Pharmaceutical Formulation and Degradation Kinetics. *J Chromatograph Separat Techniq.2014; 5:218.*
25. Abdallah NA HPLC and Densitometric TLC Methods for Simultaneous Determination of Gemifloxacin with Some Co-administered Drugs in Human Plasma. *J Chromatograph Separat Techniq. 2014; 5:220*
26. Vijaya Bhaskar V et.al. Liquid Chromatography/Tandem Mass Spectrometry Method for Quantitative Estimation of Cremophor EL and its Applications. *J Anal Bioanal Tech.2013; 4:163.*
27. Burhenne J Bioanalytical Method Validation. *J Anal Bioanal Tech.2012; 3:e111*
28. Ramani AV et al. Study of Pharmacokinetics and Tissue Distribution of BITS-17 in Rat Plasma and Tissue Homogenate Using a Validated LC Method. *J Bioanal Biomed.2013; 4:079-084.*
29. Naveed S Analytical Determination of Lisinopril Using UV Spectrophotometer and HPLC: An Overview. *Mod Chem appl.2014; 2: 137.*
30. Murthy TGK and Geethanjali J, Development of a Validated RP-HPLC Method for Simultaneous Estimation of Metformin Hydrochloride and Rosuvastatin Calcium in Bulk and In-House Formulation. *J Chromatogr Sep Tech 2014, 5: 252*
31. Pushpa Kumari K et.al. Stability Indicating RP-HPLC method Development and Validation of Salicylic Acid in Choline Magnesium Trisalicylate (Trilisate) Tablets. *J Pharma Care Health Sys 2014, 1: 4*
32. Whitmire M et.al. LC-MS/MS Bioanalysis Method Development, Validation, and Sample Analysis: Points to Consider When Conducting Nonclinical and Clinical Studies in Accordance with Current Regulatory Guidances. *J Anal Bioanal Techniques.2011; S4:001.*
33. Nouruddin W Ali et.al. Development and Validation of Different Chromatographic Methods for Determination of Two Hypouricemic Drugs in Their Combined Dosage Form. *J Anal Bioanal Tech 2014, 5:211*